2025/05/14
Distributed by FUJIFILM Wako Pure Chemical Corporation
Manufactured by FUJIFILM Wako Pure Chemical Corporation

Code No.29-3522

SPECIFICATION

BioCleanPlate WakoTM

for Endotoxin Assay

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Bacterial endotoxins	less than 0.01EU/mL
Control number	measured value

revised on 2025/02/13